

TEST REPORT: 7191238256-CHM20-03-RC
09 JUN 2020



PSB Singapore

RESULTS

Sample Description : Damon Mask – Filter M (Before wash)

Test sample/ controls	Stage 1, CFU	Stage 2, CFU	Stage 3, CFU	Stage 4, CFU	Stage 5, CFU	Stage 6, CFU	Sum of Total plate count for the 6 sieves, CFU	Average Count for Controls, CFU	BFE (%)	Average BFE (%)
-ve Control	0	0	0	0	0	0	0			
+ve Control 1	169	460	802	645	449	124	2649	2696		
+ve Control 2	183	482	778	620	489	190	2742			
Sample 1	0	0	0	0	2	35	37		98.63	98.10
Sample 2	0	0	0	0	4	47	51		98.11	
Sample 3	0	0	0	0	1	46	47		98.26	
Sample 4	0	0	0	0	9	57	66		97.55	
Sample 5	0	0	0	0	6	49	55		97.96	